

Home User App Compatibility

Not all **Android**[™] and **Apple**[®] phones are compatible with the optional myBioness[™] mobile app. L300 Go[®] and L360 are designed to be operable with or without the application in such cases. Refer to the following App store links for information on minimum system requirements.

Android phones: <https://play.google.com/store/apps/details?id=com.bioness.mapp>

Apple phones: <https://apps.apple.com/us/app/mybioness/id1037396102>

An L300 Go or L360 device is a prerequisite for using the myBioness application. The application is compatible with External Pulse Generator (EPG) firmware version 1.53 and higher. Since Android is an open platform with many variations of Bluetooth[®] technology, we recommend the following process for verifying phone compatibility: download the myBioness app to your Android phone and follow the in-app instructions to pair your EPG device(s). Refer to your phone's user guide to verify appropriate permissions and settings for general Bluetooth pairing. Although successful pairing customarily indicates full compatibility, some features may be limited due to the impact of Android's variable platform.